

# Re-training Memorandum

**To:** EPA 8141A analyst and reviewer  
**CC:** Training file  
**From:** Sharon Dehmlow  
**Date:** 04/09/12  
**Re:** Data upload and Review (M. Johnson 64443)


---

For M. Johnson #64443, results for sample AY35970 were uploaded into Labworks before removing false positives that were outside retention time windows on the confirmation column or less than 3 X signal to noise.


As an analyst, I understand that it is my responsibility to screen samples for positive detections in EnviroQuant and to remove analytes (before upload into LIMS) that do not meet criteria to be considered a positive detection.

As a data reviewer, I understand that it is my responsibility to review sample data in EnviroQuant, as listed on the Multilevel Review Sheet, in order to verify positive detections reported on the Form 1 and to screen for potential positive detections that were missed by the analyst.

  
\_\_\_\_\_  
Analyst

  
\_\_\_\_\_  
Date

  
\_\_\_\_\_  
Data Reviewer

  
\_\_\_\_\_  
Date