



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

January 22, 2024

Teresa Coley
Sprague River Water Quality Lab
5671 Sprague River Road
Chiloquin, OR 97624
TEL: (541) 827-5231
FAX

RE: 2022 Lower Klamath Project

Order No.: 24010430

Dear Teresa Coley:

Neilson Research Corporation received 8 sample(s) on 1/11/2024 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra Schmedemann
Senior Project Manager
245 S Grape St
Medford, OR 97501



Original



**NEILSON
RESEARCH
CORPORATION**

*Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com*

Case Narrative

WO#: **24010430**
Date: **1/22/2024**


CLIENT: Sprague River Water Quality Lab

Project: 2022 Lower Klamath Project

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-01
Client Sample ID: 4011001-01
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 8:10:00 AM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	0.445 J	1	0.192	0.700	mg/L		01/12/24 14:45	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	0.364 J	1	0.0989	0.500	mg/L		01/18/24 11:06	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-02
Client Sample ID 4011001-02
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/8/2024 11:20:00 AM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED TRACE METALS										
Aluminum, Dissolved	E200.7	A	0.0137 J	1	0.00895	0.0200	mg/L		01/18/24 19:00	CJS
TRACE METALS BY EPA 200.7 ICP										
Aluminum	E200.7	A	0.523	1	0.00895	0.0200	mg/L		01/16/24 19:46	CJS
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	1.91	1	0.192	0.700	mg/L		01/12/24 15:03	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	1.97	1	0.0989	0.500	mg/L		01/18/24 10:29	TJW

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-03
Client Sample ID: 4011001-03
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/8/2024 1:22:00 PM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED TRACE METALS										
Aluminum, Dissolved	E200.7	A	0.0261	1	0.00895	0.0200	mg/L		01/18/24 19:10	CJS
TRACE METALS BY EPA 200.7 ICP										
Aluminum	E200.7	A	0.405	1	0.00895	0.0200	mg/L		01/16/24 19:50	CJS
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	3.33	1	0.192	0.700	mg/L		01/12/24 15:21	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	3.12	1	0.0989	0.500	mg/L		01/18/24 9:53	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-04
Client Sample ID 4011001-04
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 10:29:00 AM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED TRACE METALS										
Aluminum, Dissolved	E200.7	A	0.0278	1	0.00895	0.0200	mg/L		01/18/24 19:13	CJS
TRACE METALS BY EPA 200.7 ICP										
Aluminum	E200.7	A	0.291	1	0.00895	0.0200	mg/L		01/16/24 19:53	CJS
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	3.90	1	0.192	0.700	mg/L		01/12/24 15:39	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	3.68	1	0.0989	0.500	mg/L		01/18/24 11:24	TJW

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-05
Client Sample ID: 4011001-05
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 11:22:00 AM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	4.24	1	0.192	0.700	mg/L		01/12/24 15:57	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	3.86	1	0.0989	0.500	mg/L		01/18/24 11:41	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-06
Client Sample ID: 4011001-06
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 12:23:00 PM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	4.08	1	0.192	0.700	mg/L		01/12/24 16:50	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	4.06	1	0.0989	0.500	mg/L		01/18/24 11:59	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-07
Client Sample ID: 4011001-07
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 4:03:00 PM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	4.66	1	0.192	0.700	mg/L		01/12/24 17:08	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	4.01	1	0.0989	0.500	mg/L		01/18/24 12:17	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 24010430
 Date Reported: 1/22/2024

CLIENT: Sprague River Water Quality Lab
Lab ID: 24010430-08
Client Sample ID: 4011001-08
Project: 2022 Lower Klamath Project
Sample Location: Comp

Collection Date: 1/9/2024 5:00:00 PM
Received Date: 1/11/2024 10:00:00 AM
Matrix: AQUEOUS

Analyses	Method	NELAP Status	Result Qual	DF	MDL	RL	Units	MCL	Date Analyzed	Analyst
DISSOLVED ORGANIC CARBON BY SM 5310 C-2014										
Organic Carbon, Dissolved	A5310C	A	4.97	1	0.192	0.700	mg/L		01/12/24 17:27	TJW
TOTAL ORGANIC CARBON SM 5310 C-2014										
Organic Carbon, Total	A5310C	A	4.78	1	0.0989	0.500	mg/L		01/18/24 12:35	TJW

QUALIFIERS

- C1 Sample container temperature is out of limit as specified at testcode
- H Holding times for preparation or analysis exceeded
- MI Recovery outside control limits due to Matrix Interference
- PL Permit Limit
- E Value above quantitation range
- J Analyte detected below quantitation limits
- ND Not Detected at the Reporting Limit

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 24010430
 22-Jan-24

Client: Sprague River Water Quality Lab
Project: 2022 Lower Klamath Project

TestCode: DOC_W

Sample ID: MB	SampType: MBLK	TestCode: DOC_W	Units: mg/L	Prep Date: 1/12/2024	RunNo: 46599
Client ID: PBW	Batch ID: R46599	TestNo: A5310C		Analysis Date: 1/12/2024	SeqNo: 762902
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual
Organic Carbon, Dissolved	ND	0.700			

Sample ID: LCS - 14323	SampType: LCS	TestCode: DOC_W	Units: mg/L	Prep Date: 1/12/2024	RunNo: 46599
Client ID: LCSW	Batch ID: R46599	TestNo: A5310C		Analysis Date: 1/12/2024	SeqNo: 762903
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual
Organic Carbon, Dissolved	3.58	0.700	3.750	0	95.6 90 110

Sample ID: 24010399-01DDUP	SampType: DUP	TestCode: DOC_W	Units: mg/L	Prep Date: 1/12/2024	RunNo: 46599
Client ID: BatchQC	Batch ID: R46599	TestNo: A5310C		Analysis Date: 1/12/2024	SeqNo: 762905
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual
Organic Carbon, Dissolved	32.5	14.0			29.61 9.16 15

Sample ID: 24010399-02DMS	SampType: MS	TestCode: DOC_W	Units: mg/L	Prep Date: 1/12/2024	RunNo: 46599
Client ID: BatchQC	Batch ID: R46599	TestNo: A5310C		Analysis Date: 1/12/2024	SeqNo: 762907
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual
Organic Carbon, Dissolved	41.8	7.00	25.00	16.34	102 85 115

Qualifiers: C1 Sample container temperature is out of limit as specified at testcode E Value above quantitation range H Holding times for preparation or analysis exceeds
 J Analyte detected below quantitation limits MI Recovery outside control limits due to Matrix Interference ND Not Detected at the Reporting Limit
 PL Permit Limit RL Reporting Detection Limit

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 24010430
 22-Jan-24

Client: Sprague River Water Quality Lab
Project: 2022 Lower Klamath Project

TestCode: ICP_200.7_W

Sample ID: MB-24001	SampType: MBLK	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/16/2024	RunNo: 46631						
Client ID: PBW	Batch ID: 24001	TestNo: E200.7	E200.7	Analysis Date: 1/16/2024	SeqNo: 763429						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Aluminum	ND	0.0200									

Sample ID: LCS-24001	SampType: LCS	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/16/2024	RunNo: 46631						
Client ID: LCSW	Batch ID: 24001	TestNo: E200.7	E200.7	Analysis Date: 1/16/2024	SeqNo: 763430						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Aluminum	1.00	0.0200	1.000	0	100	85	115				

Sample ID: 24010341-01AMS	SampType: MS	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/16/2024	RunNo: 46631						
Client ID: BatchQC	Batch ID: 24001	TestNo: E200.7	E200.7	Analysis Date: 1/16/2024	SeqNo: 763432						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Aluminum	32.8	0.0200	11.00	14.78	164	70	130				MI

Sample ID: 24010341-01AMSD	SampType: MSD	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/16/2024	RunNo: 46631						
Client ID: BatchQC	Batch ID: 24001	TestNo: E200.7	E200.7	Analysis Date: 1/16/2024	SeqNo: 763433						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Aluminum	32.2	0.0200	11.00	14.78	159	70	130	32.78	1.66	20	MI

Qualifiers: C1 Sample container temperature is out of limit as specified at testcode
 E Value above quantitation range
 H Holding times for preparation or analysis exceeds
 J Analyte detected below quantitation limits
 MI Recovery outside control limits due to Matrix Interference
 ND Not Detected at the Reporting Limit
 PL Permit Limit
 RL Reporting Detection Limit

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 24010430
 22-Jan-24

Client: Sprague River Water Quality Lab
Project: 2022 Lower Klamath Project

TestCode: ICP_200.7_W_DISS2

Sample ID: MB-24024	SampType: MBLK	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46693
Client ID: PBW	Batch ID: 24024	TestNo: E200.7	E3005	Analysis Date: 1/18/2024	SeqNo: 764638
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Aluminum, Dissolved ND 0.0200

Sample ID: LCS-24024	SampType: LCS	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46693
Client ID: LCSW	Batch ID: 24024	TestNo: E200.7	E3005	Analysis Date: 1/18/2024	SeqNo: 764639
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Aluminum, Dissolved 1.04 0.0200 1.000 0 104 85 115

Sample ID: 24010430-02DMS	SampType: MS	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46693
Client ID: 4011001-02	Batch ID: 24024	TestNo: E200.7	E3005	Analysis Date: 1/18/2024	SeqNo: 764641
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Aluminum, Dissolved 11.1 0.0200 11.00 0.01366 101 70 130

Sample ID: 24010430-02DMSD	SampType: MSD	TestCode: ICP_200.7_W	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46693
Client ID: 4011001-02	Batch ID: 24024	TestNo: E200.7	E3005	Analysis Date: 1/18/2024	SeqNo: 764642
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Aluminum, Dissolved 11.0 0.0200 11.00 0.01366 100 70 130 11.14 0.857 20

Qualifiers: C1 Sample container temperature is out of limit as specified at testcode E Value above quantitation range H Holding times for preparation or analysis exceeds
 J Analyte detected below quantitation limits MI Recovery outside control limits due to Matrix Interference ND Not Detected at the Reporting Limit
 PL Permit Limit RL Reporting Detection Limit

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 24010430
 22-Jan-24

Client: Sprague River Water Quality Lab
Project: 2022 Lower Klamath Project

TestCode: TOC_5310C

Sample ID: MB	SampType: MBLK	TestCode: TOC_5310C	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46685						
Client ID: PBW	Batch ID: R46685	TestNo: A5310C		Analysis Date: 1/18/2024	SeqNo: 764355						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Organic Carbon, Total	0.106	0.500									J

Sample ID: LCS - 14323	SampType: LCS	TestCode: TOC_5310C	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46685						
Client ID: LCSW	Batch ID: R46685	TestNo: A5310C		Analysis Date: 1/18/2024	SeqNo: 764356						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Organic Carbon, Total	3.67	0.500	3.750	0	98.0	90	110				

Sample ID: 24010430-03ADUP	SampType: DUP	TestCode: TOC_5310C	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46685						
Client ID: 4011001-03	Batch ID: R46685	TestNo: A5310C		Analysis Date: 1/18/2024	SeqNo: 764359						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Organic Carbon, Total	3.23	0.500						3.118	3.66	15	

Sample ID: 24010430-02AMS	SampType: MS	TestCode: TOC_5310C	Units: mg/L	Prep Date: 1/18/2024	RunNo: 46685						
Client ID: 4011001-02	Batch ID: R46685	TestNo: A5310C		Analysis Date: 1/18/2024	SeqNo: 764361						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Organic Carbon, Total	4.53	0.500	2.500	1.966	102	85	115				

Qualifiers:

C1	Sample container temperature is out of limit as specified at testcode	E	Value above quantitation range	H	Holding times for preparation or analysis exceeds
J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit	RL	Reporting Detection Limit		

Original



Sample Log-In Check List

Client Name: **SPRAGUERIVERWATER**

Work Order Number: **24010430**

RcptNo: **1**

Logged by:	Ashley Spiegelberg	1/11/2024 10:00:00 AM	<i>Am</i>
Completed By:	Danielle Garten	1/11/2024 4:48:10 PM	<i>Danielle Garten</i>
Reviewed By:	Tamra Schmedemann	1/22/2024 12:51:59 PM	<i>Tamra Schmedemann</i>

Chain of Custody

1. Is Chain of Custody complete? Yes No Not Present
 2. How was the sample delivered? UPS

Log In

3. Coolers are present? Yes No NA
 4. Shipping container/cooler in good condition? Yes No
 Custody seals intact on shipping container/cooler? Yes No Not Present
 No. Seal Date: Signed By:
 5. Was an attempt made to cool the samples? Yes No NA
 6. Were all samples received at a temperature of >0° C to 6.0°C Yes No NA
 7. Sample(s) in proper container(s)? Yes No
 8. Sufficient sample volume for indicated test(s)? Yes No
 9. Are samples (except VOA and ONG) properly preserved? Yes No
 10. Was preservative added to bottles? Yes No NA
 11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes No HNO3 pH<2
 No VOA Vials
 12. Were any sample containers received broken? Yes No
 13. Does paperwork match bottle labels? Yes No
 (Note discrepancies on chain of custody)
 14. Are matrices correctly identified on Chain of Custody? Yes No
 15. Is it clear what analyses were requested? Yes No
 16. Were all holding times able to be met? Yes No
 (If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes No NA

Person Notified:	<input type="text"/>	Date:	<input type="text"/>
By Whom:	<input type="text"/>	Via:	<input type="checkbox"/> eMail <input type="checkbox"/> Phone <input type="checkbox"/> Fax <input type="checkbox"/> In Person
Regarding:	<input type="text"/>		
Client Instructions:	<input type="text"/>		

18. Additional remarks:

Cooler Information

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
1	2.5	Good				EH

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Section A Required Client Information Company: Sprague River Water Quality Lab Address: 5671 Sprague River Road Chiloquin, OR 97624 Email: teresa.coley@klamathtribes.com Phone: 541-827-5231 Fax: Collected By (Print): Ben A. Harris Collected By (Sign): <i>Ben A. Harris</i> Email Report <input checked="" type="checkbox"/> Mail Report <input type="checkbox"/> Fax Report <input type="checkbox"/>	Section B Required Project Information Project Name: Lower Klamath Project Project Number: 2022 Report To: Teresa Coley Copy To: Ben A. Harris	Section C Invoice Information Attention: Kaneeta Kirk Company Name: The Klamath Tribes Address: PO Box 436 Chiloquin, OR 97624 P.O. #	Section D Rush Status (Subject to Scheduling) <input checked="" type="checkbox"/> Standard: 10 Business Days <input type="checkbox"/> Priority: 5 Business Days (List x 1.50) <input type="checkbox"/> Express: 3 Business Days (List x 1.75) <input type="checkbox"/> Rush: 2 Business Days (List x 2.00) <input type="checkbox"/> Rush: 1 Business Day (List x 2.50) <input type="checkbox"/> Rush: Same Day (List x 3.00) Authorized <input type="checkbox"/> Yes <input type="checkbox"/> No
----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Section E Sample Information					Analysis Requested										NRC Workorder # <i>24010430</i> (Lab Use Only)				
Sample ID	Comp/Grab	Matrix*	Date Collected	Time Collected	No. of Containers	TOC	DOC	Total Aluminum	Dissolved Aluminum								Remarks / Field Data	NRC Sample # <small>(Lab Use Only)</small>	(Lab)
4011001-01	comp	water	1-9-2024	0810	6	✓	✓										please	<i>01</i>	
4011001-02	comp	water	1-8-2024	1120	8	✓	✓	✓	✓								return	<i>02</i>	
4011001-03	comp	water	1-8-2024	1322	8	✓	✓	✓	✓								ice	<i>03</i>	
4011001-04	comp	water	1-9-2024	1029	8	✓	✓	✓	✓								bottles	<i>04</i>	
4011001-05	comp	water	1-9-2024	1122	6	✓	✓										and	<i>05</i>	
4011001-06	comp	water	1-9-2024	1223	6	✓	✓										cooler	<i>06</i>	
4011001-07	comp	water	1-9-2024	1603	6	✓	✓											<i>07</i>	
4011001-08	comp	water	1-9-2024	1700	6	✓	✓											<i>08</i>	

*Matrix: DW - Drinking Water WW - Wastewater W - Water S - Soil/Solid SL - Sludge O - Oil WP - Wipe OT - Other

Section F		Sign	Print	Date	Time
Relinquished By:					
Received By:	<i>Ben A. Harris</i>	<i>Ben A. Harris</i>		<i>1-10-24</i>	<i>0841</i>
Relinquished By:					
Received By:					
Relinquished By:	<i>Ben A. Harris</i>	<i>Ben A. Harris</i>		<i>1-10-24</i>	<i>0841</i>
Received By Laboratory:	<i>Ashley Spiegelberg</i>	<i>Ashley Spiegelberg</i>		<i>1/11/24</i>	<i>10:00</i>

Section G	
Temp:	<i>2.5 IR-5</i>
≤6°C:	<input checked="" type="radio"/> Yes <input type="radio"/> No
Received on Ice:	<input type="radio"/> Yes <input checked="" type="radio"/> No
Number of Bottles Received:	
pH Checked:	<input checked="" type="checkbox"/>
COC Seals Intact:	<input type="radio"/> Yes <input type="radio"/> No <input checked="" type="radio"/> NA
Field Blank Included:	<input type="radio"/> Yes <input checked="" type="radio"/> No
Received Via:	<input checked="" type="radio"/> UPS <input type="radio"/> FedEx <input type="radio"/> Other <input type="radio"/> Hand

Payment:	<input checked="" type="radio"/> Invoice <input type="radio"/> Cash <input type="radio"/> VISA, M/C <input type="radio"/> Check # <input type="text"/> Amount <input type="text"/>
----------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

-
- A Total Alkalinity and Bicarbonate Alkalinity results are to a pH endpoint of 4.5. Carbonate Alkalinity result is to a pH endpoint of 8.3.
- A-LL The total low level alkalinity results are to a pH endpoint of 4.3-4.7 pH units per SM 2320B-2011.
- B Analyte detected in the associated method blank.
- C Sample(s) does not meet NELAP/ORELAP sample acceptance criteria. See Case Narrative.
- C1 Sample(s) does not meet NELAP/ORELAP sample acceptance criteria for temperature.
- CF Results confirmed by re-analysis.
- CU Cleanup performed as specified by method.
- E Estimated value.
- ER Elevated reporting limit due to matrix. Report limits (MDLs, MRLs & PQLs) are adjusted based on variations in sample preparation amounts, analytical dilutions, and percent solids, where applicable.
- FC Fecal Coliforms: Sample(s) received past 40 CFR Part 136 specified holding time. Results reported as estimated values.
- HP Sample re-analysis performed outside of method specified holding time.
- HR Sample received outside of method specified holding time.
- HS Sample analyzed for volatile organics contained headspace.
- HT At the client's request, the sample was analyzed outside of method specified holding time.
- H Analysis performed outside of method specified holding time.
- J Analyte detected below the Minimum Reporting Limit (MRL) and above the Method Detection Limit (MDL). The J flag result is an estimated value and the user should be aware that this data is of limited reliability.
- L Dissolved metals were not filtered within 15 minutes of collection per 40 CFR Part 136.
- MI Surrogate, Duplicate Sample (DUP) or Matrix Spikes recoveries are out of control limits due to matrix interference. Sample results may be biased.
- N See Case Narrative on page 2 of report.
- Q Initial calibration verification (ICV), continuing calibration verification (CCV) or laboratory control sample (LCS), and/or matrix spikes exceeded high recovery limits, but associated samples are non-detect and the sample results are not affected. Data meets EPA/NELAP requirements.
- R Relative percent difference (RPD) is outside of the accepted recovery limits.
- R3 The relative percent difference (RPD) and/or percent recovery for the duplicate (DUP) or matrix spike (MS)/matrix spike duplicate (MSD) cannot be accurately calculated due to the concentration of analyte already present in the sample.
- R4 The Relative percent difference (RPD) is not within control limits because the concentration of the sample result is too low to represent proper statistical error.
- R5 The difference between the BOD/CBOD results for the highest and lowest dilution used for the calculation is >30% because the results are too low to represent proper statistical error. The BOD/CBOD sample result is an average of all qualified bottles for each dilution series. The sample results are not affected.
- R6 The difference between the BOD/CBOD results for the highest and lowest dilution used for the calculation is >30%. This may indicate a possible matrix interference. The BOD/CBOD sample result is an average of all qualified bottles for each dilution series.
- S Surrogate and/or matrix spike recovery is outside of the accepted recovery limits. Sample results may be biased.
- S1 Surrogate or matrix spike recovery is outside of control limits due to dilution necessary for analysis.
- SC Sub-contracted to another laboratory for analysis.
- SP Sample(s) were not collected per EPA Method 5035A protocols. The results are considered minimum values.
- * Value exceeds Maximum Contaminant Level or is outside the acceptable range.